



Recommended Maximum Pacing Duration for Defibrillator Pads (In Hours)

20ms Pulse Width Example: **Philips** Codemaster

| | | Pulse Amplitude (mA) | | | | | | | | | | | | | | | | | | | |
|-------------------|-----|----------------------|----|----|----|----|----|----|----|----|-----|-----|-----|-----|-----|-----|-----|-----|-----|-----|-----|
| | | 10 | 20 | 30 | 40 | 50 | 60 | 70 | 80 | 90 | 100 | 110 | 120 | 130 | 140 | 150 | 160 | 170 | 180 | 190 | 200 |
| Pacing Rate (ppm) | 40 | 24 | 24 | 24 | 24 | 24 | 24 | 24 | 22 | 20 | 18 | 18 | 16 | 14 | 12 | 11 | 10 | 8 | 8 | | |
| | 50 | 24 | 24 | 24 | 24 | 24 | 22 | 23 | 21 | 20 | 18 | 17 | 16 | 14 | 12 | 10 | 8 | 8 | 7 | | |
| | 60 | 24 | 24 | 24 | 24 | 22 | 20 | 20 | 20 | 18 | 17 | 16 | 15 | 13 | 11 | 8 | 6 | 4 | 4 | | |
| | 70 | 24 | 24 | 24 | 22 | 20 | 18 | 17 | 17 | 16 | 16 | 15 | 13 | 11 | 10 | 6 | 5 | 3 | 3 | | |
| | 80 | 24 | 24 | 24 | 22 | 18 | 16 | 15 | 14 | 14 | 13 | 12 | 11 | 10 | 9 | 6 | 4 | 3 | 3 | | |
| | 90 | 24 | 24 | 23 | 18 | 16 | 14 | 14 | 13 | 13 | 12 | 10 | 10 | 9 | 8 | 5 | 3 | 3 | 2 | | |
| | 100 | 24 | 24 | 20 | 18 | 16 | 14 | 12 | 12 | 11 | 10 | 9 | 8 | 8 | 7 | 5 | 3 | 2 | 2 | | |
| | 110 | 24 | 24 | 16 | 16 | 12 | 11 | 11 | 11 | 9 | 9 | 8 | 8 | 7 | 6 | 4 | 3 | 2 | 2 | | |
| | 120 | 24 | 24 | 16 | 14 | 12 | 10 | 9 | 8 | 8 | 7 | 6 | 5 | 5 | 4 | 4 | 2 | 2 | 2 | | |
| | 130 | 24 | 24 | 14 | 14 | 10 | 9 | 8 | 7 | 6 | 5 | 4 | 4 | 4 | 4 | 3 | 2 | 2 | 1 | | |
| | 140 | 24 | 24 | 12 | 10 | 10 | 8 | 6 | 5 | 4 | 4 | 4 | 3 | 3 | 3 | 2 | 2 | 2 | 1 | | |
| | 150 | 24 | 24 | 10 | 10 | 8 | 6 | 5 | 4 | 4 | 4 | 3 | 3 | 2 | 2 | 2 | 2 | 1 | 1 | | |
| | 160 | 24 | 24 | 8 | 8 | 8 | 5 | 5 | 4 | 3 | 3 | 3 | 2 | 2 | 2 | 2 | 1 | 1 | 1 | | |
| 170 | 24 | 24 | 8 | 8 | 6 | 5 | 4 | 4 | 3 | 3 | 3 | 2 | 2 | 2 | 2 | 1 | 1 | 1 | | | |
| 180 | 24 | 24 | 8 | 8 | 6 | 5 | 4 | 3 | 2 | 2 | 2 | 1 | 1 | 1 | 1 | 1 | 1 | 1 | | | |